



Quaternary ammonium compounds leave non-volatile residues that must be rinsed with sterile water or alcohol before procedure setup.

CLEAN IT ONCE TO DISINFECT, RINSE IT TWICE TO PROTECT.