



Verify autoclave functionality with spore tests.

Confirm single-use needle disposal records.

Check ink bottle expiration dates.

Ensure cross-contamination prevention protocols.

Routine inspections verify critical sterilization and cross-contamination protocols.

Focus on verifiable risks.