Fill with distilled water.

Pre-clean and dry instruments.

Pouch items individually.

Run vacuum cycle.

Organic material on an instrument shields microorganisms from steam.

The autoclave does not clean.